

User Instruction for the Extra Wide High Back Chair – Brown X162



1. Introduction

Your High Back Chair is made with easy clean upholstery for longevity of service. The tubular metal framework is coated in high gloss nylon epoxy powder coating. The chair is adjustable in height and lumbar support is given with the profiled back. The framework design and none-slip rubber feet both work towards patient stability; and the two transportation wheels ease mobility. (Please note the transport wheels are designed to aid the movement of the product, **this is not a mobile commode!**)

Product Dimensions	
Seat height (from-to)	470mm-572mm
Width between arms	585mm
Seat width	610mm
Seat depth	508mm
Footprint	787mm x 648mm

2. Contents of the Box

- Bariatric High Back Chair – brown (X162)
- Extended Legs & Pin Clips
- Transfer Wheels
- User Instructions (this leaflet)

3. Assembly Instructions

- Carefully unpack and examine the product to ensure it is complete.
- With the chair standing on its four feet, remove the pin clip from the left front leg and rotate the leg with the wheel so that the wheel faces out from the chair.
- Replace the pin clip to secure the leg in place
- Repeat those steps on the rear left leg. The transportation wheels are now ready for use.
- Before use check for any signs of damage e.g. cracked weld, flaking of paint, bent tube. If any signs occur please contact your supplier to exchange and refrain from using the product. **DO NOT ATTEMPT TO USE IF PRODUCT IS DAMAGED.**

User Instruction for the Extra Wide High Back Chair – Brown X162

4. Correct Use

Adjustable Height:

- Remove the pin clips from the legs
- Select the desired height of the frame by gently sliding out each extension leg from the frame (NOTE – this is a telescopic action).
- At desired height refit the pin clips, ensuring its passes through the extension leg fully. Adjust all other legs in the same manner, ensuring that they are all the same height.

5. Cleaning

- Use mild detergent with warm water to wipe down framework on at least a weekly basis more frequently if required.
- Do not use an abrasive cleaner as this will damage the frame and nullify your warranty.
- This product is not suitable to be cleaned using power/jet washing & industrial washing machines.

6. Maintenance Requirements

The product should be checked on a regular basis for damage or wear and tear – this includes flaking or cracking paint. Any undue signs of damage should be noted and reported to the supplier in the first instance. Should damage be evident, then the user should refrain from using the product until advice has been sought from the supplier.

7. Warning

- This product is intended for users of a maximum of 250kg/40stone in weight.
- We recommend a full risk assessment should be carried out before use.

8. Warranty

This product carries a 12 month warranty from the date of purchase against manufacturing defects. Misuse or damage is not covered. This warranty does not affect your statutory rights. As manufacturers we cannot specify expected product life as this is based on the level of usage.

9. Disposal Instructions

- Appropriate hygiene standards should be applied before disposal to reduce the risk of infection.
- Materials used are essentially non hazardous and can be disposed of appropriately.

10. Contact Details

We have attempted to cover most eventualities, however if you have any specific queries, please contact:

Cefndy Healthcare
Cefndy Road
Rhyl
Denbighshire
LL18 2HG
Tel: 0044 1745 343877 (Quality Assurance Dept)
E-mail: cefndyengineering@denbighshire.gov.uk

